INTOX EC/IR II

Quality Assurance Worksheet

010528 Instrument Serial Number

Worksheet Start Date

12/4/2014

Fairfax County Adult Detention Center Location

Address 10520 Judicial Dr., Fairfax, VA 22030

18910 License No. Heather Stanton DFS Technician

> T Laboratory ∇ On-Site

Site Specification: No detrimental environmental conditions exist. V

760 Reference Barometer (mm HG) 761 Instrument Barometer (mm HG)

4/28/2015 009113 **RB** Calibration Due Reference Barometer(RB)Serial #

Measurement Assurance Check

Sample 1 Standard (sea level) maximum **PA Target** minimum 0.003 Sample 2 0.300 -0.003

Sample 3 Precision sample min sample max

0.000 0.000 0

Standard (sea level) **PA Target** minimum maximum Sample 1 -0.003 0.003 Sample 2 0.100

sample min sample max Sample 3 Precision 0 0.000 0.000

AG419702-50 Dry gas standard Lot No. (with tank no.)

Supplies Replaced dry gas standard (+O-ring)

Mouthpieces Installed at Location Certificates of Analysis

Operator Worksheet □ Removed to DFS-Central Other:

Notes:

Calibrated	Certified	Г Meas	urement Assurance	Check
Instrument Test uccessful completion denotes	satisfactory conditio		leshooting/Maintenan	ce Worksheet Complet
Standard (sea level 0.29 Precision 0		minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level 0.18 Precision 0		minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level 0.08 Precision 0		minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level 0.02 Precision 0	· —	minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3

R II QA Worksheet

Issuing Analyst

Intox EC/IR II QA Worksheet Issued by Calibration and Training Program Manager Issue Date: 9-December-2013 DFS Document 250-F113 Revision 7 Page 2 of 2

Date





COMMONWEALTH OF VIRGINIA DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED		NAME OF COURT	
INSTRUMENT, TEST,		DFS	
BREATH ANALYSIS			
SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY	
STANTON, HEATHER, M		DFS Central Lab	
DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED	
18910	10/01/2016	12/04/2014	
TEST EQUIPMENT NUMBER		於	
010528			
RESULTS: TIME SAMPLE TAKEN	10:18 EST		
SAMPLE'S ALCOHOL CONTENT	0.00	GRAMS PER 210 LITERS OF BREATH	
ATTIE OT.			
ATTEST: I CERTIFY THAT THE ABOVE IS AN ACCURATE	E RECORD OF THE TEST	CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF	
EQUIPMENT AND IN ACCORDANCE WITH TH CONDUCTED IN ACCORDANCE WITH THE TEST THE ACCUSED WAS ADVISED OF HIS	IE METHODS APPROVED THE DEPARTMENT'S S RIGHT TO OBSERVE TH AND THAT I POSSESS A	D BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF HE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIP A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND	
BREATH TEST OPERATOR	(0.4.4.600.0.3.4.4.4.4.4.4.4.4.4.4.4.4.4.4.4.4.4		
☐ I HAVE RECEIVED A COPY OF THIS	S CERTIFICATE OF	ANALYSIS	
		SUBJECT'S SIGNATURE	
☐ SUBJECT REFUSED TO SIGN FOR C	COPY OF CERTIFICA	TE OF ANALYSIS	
		OPERATOR'S SIGNATURE	



IntoxNet MIS Report Report Generated 04 Dec 2014 at 18:43

Test Results

	Instrument Serial Number 010	528
Test # 001398 Subject Test		
Test Location 1 Department of	Test Location 2 Forensic Science	Test Location 3
Test Date 04 Dec 2014	Test Time 10:11	System Check Passed
	Remote/Local Local	
Operator's Last Name STANTON	Operator's First Name HEATHER	Operator's Middle Initial V
Agency DFS Central Lab	License	Number 18910
Card Serial Number 118910	Effective Date 10/01/2014	Expiration Date 10/01/2016
Subject's Last Name INSTRUMENT	Subject's	First Name TEST
Subject's Middle Initial	Subject's Date of Birth 00/00/0000	Subject's Sex Male
Driver's License Number	Dr	ver's License Expiration 00/00/0000
Driver's License State	Court Na	ne DFS
End Date 04 Dec 2014 End Time 10	0:19 Result Time 10:18 R	esult Date 04 Dec 2014 Result 0.00
Data Type DIAG	Sample Value Pass	Sample Time 10:12
Data Type BLK	Sample Value 0.000	Sample Time 10:13
Data Type CHK	Sample Value 0.101	Sample Time 10:13
Data Type BLK	Sample Value 0.000	Sample Time 10:15
Data Type SUBJ	Sample Value 0.000	Sample Time 10:15
Data Type BLK	Sample Value 0.000	Sample Time 10:16
Data Type SUBJ	Sample Value 0.000	Sample Time 10:18
Data Type BLK	Sample Value 0.000	Sample Time 10:19
Standard Type Dry Gas Std		Standard Value 0.100
Standard Lot Number AG419702-50		tandard Expiration Date 07/16/2016
Tank Pressure 1125	Barom	etric Pressure 760 mmHg
Blow Sample Number 1 Blow Du		1751 cc End-of-Blow Time 10:15
Blow Sample Number 2 Blow Du	ration 3.38 sec Blow Volume	1773 cc

Tamper Evident Stamp 383c8624

Test Status Success

Test Status Code 0